



TestPro Firmware Update using USB Drive


This article describes how to update the Firmware of TestPro using a USB Drive.

1. Connect USB Drive to PC


Ensure that the USB Drive is in FAT32 format and there is at least 200MB of free space on drive.

2. Open TestDataPro PC Software

Click on "Create USB Upgrade".



TestDataPro will download the latest Software Upgrade Package, Unzip & copy to USB Drive.



3. Update TestPro

Unplug the USB Drive from PC.

Turn OFF TestPro.

Connect USB Drive to TestPro & Power ON TestPro.

Note : Keep Power Adapter connected to TestPro at all times during upgrade.

Once update is complete, the Display will prompt user to remove the USB drive.

Both Main & Remote Units must be updated with the same software Version for proper functionality.

4. Check Software Version

TestPro Main Menu => Settings => Device Information